

Healthcare Sector

Certification of Compatibility

Trailer Manufacturer:	Kentucky Trailer Technologies		
	1240 N. Pontiac Trail		
	Walled Lake, MI 48390, U.S.A.		
Trailer Type:	Trailer Siemens 1.5T		
MRI Systems:	MAGNETOM Avanto & MAGNETOM Es	spree	
Test Results against SIEMEN	S SPECIFICATIONS	YES	NO
 SIEMENS suspension specification fulfilled and documented Stray field shielding specification fulfilled and documented Written certification that air conditioners, water chiller and humidifier system will perform as specified Manuals and documentation for all installed equipment and supplied equipment Detailed drawings which show all equipment locations and intended routes for all SIEMENS supplied cables Written certification that RF room meets the specified requirements Detailed site planning guide Detailed preventive maintenance schedule 			000 0 0 000
We hereby issue to the manufacturer a certification of compatibility for the above mentioned trailer or relocatable for the combined use with the above mentioned Siemens MRI unit.			
Erlangen, March 10, 2015	*	1	
pa. Ball	i.V. 28	~	•••
Dr. Bernd Ohnesorge, CEO MR Dr. Reiner Bachus, VP Quality MR			1R

Siemens Aktiengesellschaft

DISCLAIMER: THE CERTIFICATION OF COMPATIBILITY SHALL BECOME INVALID AS SOON AS THE MANUFACTURER MAKES ANY CHANGES TO HIS PRODUCTS THAT DO NOT MEET THE SPECIFICATIONS. THE CERTIFICATION OF COMPATIBILITY CAN ALSO CEASE TO BE VALID IF THE MANUFACTURER BECOMES AWARE OF CUSTOMER COMPLAINTS WHICH CAN BE ATTRIBUTED TO COMBINED USE. THE MANUFACTURER SHALL BE OBLIGED TO INFORM SIEMENS WITHOUT DELAY WHENEVER HE BECOMES AWARE OF ANY SUCH COMPLAINTS, OR WHENEVER HE MAKES ANY CHANGES TO HIS PRODUCT THAT DO NOT MEET THE SPECIFICATIONS. THE MANUFACTURER SHALL BE OBLIGED TO SOLVE ANY CUSTOMER COMPLAINTS REGARDING HIS PROD-UCTS. SIEMENS SHALL BE ENTITLED TO REVOKE THE CERTIFICATION OF COMPATIBILITY, GIVING THE REASONS FOR DOING SO. POSSIBLE GROUNDS FOR DOING SO SHALL INCLUDE, BUT NOT BE LIMITED TO THE FOLLOWING: NEW COMPATIBILITY PROBLEMS WHICH HAVE ARISEN, CHANGES TO ANY OF THE MRI SYSTEMS NAMED IN THE CERTIFICATION OF COMPATIBILITY, CUSTOMER COMPLAINTS IN CONNECTION WITH THE COMBINED USE. ALL COMMUNICATIONS BETWEEN THE MANUFACTURER AND SIEMENS SHALL BE MADE IN WRITING AND WITHOUT DELAY; ORAL (INCLUDING TELEPHONE) COMMUNICATIONS SHALL BE CONFIRMED IN WRITING WITHOUT DELAY. REGARDLESS OF WHETHER AND WHEN SIEMENS IS NOTIFIED, THE VALIDITY OF A COMPATIBILITY CERTIFICATE SHALL END ON THE DAY ON WHICH THE MANUFACTURER BECOMES AWARE OF A CUSTOMER COMPLAINT IN CONNECTION WITH THE COMBINED USE, MAKES CHANGES TO HIS PRODUCTS THAT DO NOT MEET THE SPECI-FICATIONS, OR RECEIVES A REVOCATION FROM SIEMENS. THE MANUFACTURER SHALL INFORM HIS CUSTOMERS PROMPTLY OF THE CANCELLATION OF VALIDITY OF THIS CERTIFICATION OF COMPATIBILITY.