



Certification of Compatibility

Trailer Manufacturer: Advanced Mobility by Kentucky Trailer
6370 W. Emerald Parkway, Suite 106
Monee, IL 60449, USA

MRI System: MAGNETOM Aera with XJ gradient configuration

Trailer Type: US Trailer

Trailer Design: MAGNETOM Aera shield design

Test Results against SIEMENS SPECIFICATIONS	YES	NO
• SIEMENS suspension specification fulfilled and documented	<input checked="" type="checkbox"/>	<input type="checkbox"/>
• Stray field shielding specification fulfilled and documented	<input checked="" type="checkbox"/>	<input type="checkbox"/>
• Written certification that air conditioners, water chiller and humidifier system will perform as specified	<input checked="" type="checkbox"/>	<input type="checkbox"/>
• Manuals and documentation for all installed equipment and supplied equipment	<input checked="" type="checkbox"/>	<input type="checkbox"/>
• Detailed drawings which show all equipment locations and intended routes for all SIEMENS supplied cables	<input checked="" type="checkbox"/>	<input type="checkbox"/>
• Written certification that RF room meets the specified requirements	<input checked="" type="checkbox"/>	<input type="checkbox"/>
• Detailed site planning guide	<input checked="" type="checkbox"/>	<input type="checkbox"/>
• Detailed preventive maintenance schedule	<input checked="" type="checkbox"/>	<input type="checkbox"/>

We hereby issue to the manufacturer a certification of compatibility for the above mentioned trailer for the combined use with the above mentioned Siemens MRI unit.

Erlangen, November 14, 2018


.....
Dr. Arthur Kaindl
Head of Business Line MR


.....
Reiner Lochner
Head of Quality&Technologies

Siemens Healthineers

DISCLAIMER: THE CERTIFICATION OF COMPATIBILITY SHALL BECOME INVALID AS SOON AS THE MANUFACTURER MAKES ANY CHANGES TO HIS PRODUCTS THAT DO NOT MEET THE SPECIFICATIONS. THE CERTIFICATION OF COMPATIBILITY CAN ALSO CEASE TO BE VALID IF THE MANUFACTURER BECOMES AWARE OF CUSTOMER COMPLAINTS WHICH CAN BE ATTRIBUTED TO COMBINED USE. THE MANUFACTURER SHALL BE OBLIGED TO INFORM SIEMENS WITHOUT DELAY WHENEVER HE BECOMES AWARE OF ANY SUCH COMPLAINTS, OR WHENEVER HE MAKES ANY CHANGES TO HIS PRODUCT THAT DO NOT MEET THE SPECIFICATIONS. THE MANUFACTURER SHALL BE OBLIGED TO SOLVE ANY CUSTOMER COMPLAINTS REGARDING HIS PRODUCTS. SIEMENS SHALL BE ENTITLED TO REVOKE THE CERTIFICATION OF COMPATIBILITY, GIVING THE REASONS FOR DOING SO. POSSIBLE GROUNDS FOR DOING SO SHALL INCLUDE, BUT NOT BE LIMITED TO THE FOLLOWING: NEW COMPATIBILITY PROBLEMS WHICH HAVE ARISEN, CHANGES TO ANY OF THE MRI SYSTEMS NAMED IN THE CERTIFICATION OF COMPATIBILITY, CUSTOMER COMPLAINTS IN CONNECTION WITH THE COMBINED USE. ALL COMMUNICATIONS BETWEEN THE MANUFACTURER AND SIEMENS SHALL BE MADE IN WRITING AND WITHOUT DELAY; ORAL (INCLUDING TELEPHONE) COMMUNICATIONS SHALL BE CONFIRMED IN WRITING WITHOUT DELAY. REGARDLESS OF WHETHER AND WHEN SIEMENS IS NOTIFIED, THE VALIDITY OF A COMPATIBILITY CERTIFICATE SHALL END ON THE DAY ON WHICH THE MANUFACTURER BECOMES AWARE OF A CUSTOMER COMPLAINT IN CONNECTION WITH THE COMBINED USE, MAKES CHANGES TO HIS PRODUCTS THAT DO NOT MEET THE SPECIFICATIONS, OR RECEIVES A REVOCATION FROM SIEMENS. THE MANUFACTURER SHALL INFORM HIS CUSTOMERS PROMPTLY OF THE CANCELLATION OF VALIDITY OF THIS CERTIFICATION OF COMPATIBILITY.